

Table of contents

\\USER

Brain

HARP_Prisma_XA60A

HARP_20190617

AAHead_Scout
 AC-PC_setup
 SEField1_AP
 BOLD_REST1_AP
 SEField1_PA
 BOLD_REST1_PA
 T1_MPR
 T2_SPC
 ASL_ADNI
 QSM_3D
 DWI_AP
 DWI_PA
 SEField2_AP
 BOLD_REST2_AP
 SEField2_PA
 BOLD_REST2_PA
 SEField3_AP
 BOLD_REST3_AP
 SEField3_PA
 BOLD_REST3_PA
 SEField4_AP
 SEField4_PA
 BOLD_EMOTION_PA
 SEField5_AP
 SEField5_PA
 BOLD_CARIT_PA

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\AAHead_Scout

TA: 14 sec Coil Selection: Auto Voxel Size: 1.6×1.6×1.6 mm³ Acc:: 3 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	On
Graphic segment	Default
Inline Movie	Off

Resolution - Common

Slice Thickness	1.6 mm
Base Resolution	160
Phase Resolution	100 %
Slice Resolution	69 %
Trajectory	Cartesian

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	3
Reference Lines PE	24
Acceleration Factor 3D	1
Phase Partial Fourier	6/8
Slice Partial Fourier	6/8
Asymmetric Echo	Weak

Routine

Slab Group	1
Slabs	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	128
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FOV Read	260 mm
FOV Phase	100.0 %
Slice Thickness	1.6 mm
TR	3.15 ms
TE	1.37 ms
Averages	1
Concatenations	1
AutoAlign	Head

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	2D
Normalize	Prescan
Noise Masking	Off
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	128
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FOV Read	260 mm
FOV Phase	100.0 %
Slice Thickness	1.6 mm
TR	3.15 ms
Multi-Slice Mode	Sequential
Series	Ascending
Concatenations	1

Contrast - Common

TR	3.15 ms
TE	1.37 ms
Flip Angle	8 deg
Fat-Water Contrast	Standard
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Time to Center	6.2 s

Resolution - Common

FOV Read	260 mm
FOV Phase	100.0 %

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head
Initial Position	Isocenter
L	0.0 mm

Geometry - AutoAlign

P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Inline - MapIt

MapIt	None
Flip Angle	8 deg
Measurements	1
Contrasts	1
TE	1.37 ms
TR	3.15 ms
Save Original Images	On

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	fl
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Fast
Gradient Mode	Normal
Bandwidth	540 Hz/Px
Asymmetric Echo	Weak

Sequence - Part 2

Introduction	On
RF Spoiling	On
Breast Application	Off
Phase Enc. Order	Automatic

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\AC-PC_setup

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\SEField1_AP

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\BOLD_REST1_AP

TA: 5:08 min Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	34.40 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	Head > Brain

Contrast - Common

TR	800.0 ms
TE	34.40 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	375
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	375
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Flow Compensation	None
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	Off
EPI Factor	86

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg

Sequence - Special

Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\SEField1_PA

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\BOLD_REST1_PA

TA: 5:08 min Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	34.40 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	Head > Brain

Contrast - Common

TR	800.0 ms
TE	34.40 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	375
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	375
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Flow Compensation	None
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	Off
EPI Factor	86

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg

Sequence - Special

Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\T1_MPR

TA: 5:22 min Coil Selection: Auto Voxel Size: 0.8×0.8×0.8 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	On
Graphic segment	1st Segment
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	224
Phase Oversampling	0 %
Slice Oversampling	7.1 %
FOV Read	256 mm
FOV Phase	93.8 %
Slice Thickness	0.80 mm
TR	2500.0 ms
TE	2.18 ms
Averages	1
Concatenations	1
AutoAlign	Head > Brain

Contrast - Common

TR	2500.0 ms
TE	2.18 ms
Magn. Preparation	Non-sel. IR
TI	1000 ms
Flip Angle	8 deg
Fat-Water Contrast	Fast Water Excitation
Dark Blood	Off
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement
Reordering	Linear

Resolution - Common

FOV Read	256 mm
FOV Phase	93.8 %
Slice Thickness	0.80 mm
Base Resolution	320
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	32
Acceleration Factor 3D	1
Phase Partial Fourier	6/8
Slice Partial Fourier	Off
Asymmetric Echo	Allowed
Elliptical Scanning	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	224
Phase Oversampling	0 %
Slice Oversampling	7.1 %
FOV Read	256 mm
FOV Phase	93.8 %
Slice Thickness	0.80 mm
TR	2500.0 ms
Multi-Slice Mode	Single Shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain

Geometry - AutoAlign

Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	-0.80 deg

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	-0.80 deg
A >> P	240 mm
F >> H	256 mm
R >> L	180 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	5.000

Physio - Signal

1st Signal/Mode	None
TR	2500.0 ms
Concatenations	1

Physio - Cardiac

Fat-Water Contrast	Fast Water Excitation
Magn. Preparation	Non-sel. IR
TI	1000 ms
Dark Blood	Off
FOV Read	256 mm
FOV Phase	93.8 %
Phase Resolution	100 %

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Inline - MapIt

MapIt	None
Flip Angle	8 deg
Measurements	1
Contrasts	1
TE	2.18 ms
TR	2500.0 ms
Save Original Images	On

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	tfl
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Fast

Sequence - Part 1

Gradient Mode	Fast
Flow Compensation	None
Reordering	Linear
Bandwidth	220 Hz/Px
Echo Spacing	7.92 ms
Asymmetric Echo	Allowed
Turbo Factor	240

Sequence - Part 2

Introduction	On
RF Spoiling	On
Incr. Gradient Spoiling	Off

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\T2_SPC

TA: 5:34 min Coil Selection: Auto Voxel Size: 0.8x0.8x0.8 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	224
Phase Oversampling	0 %
Slice Oversampling	7.1 %
FOV Read	256 mm
FOV Phase	93.8 %
Slice Thickness	0.80 mm
TR	3200.0 ms
TE	564.00 ms
Averages	1.0
Concatenations	1
AutoAlign	Head > Brain

Contrast - Common

TR	3200.0 ms
TE	564.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle Mode	T2 Var
Fat-Water Contrast	Standard
Dark Blood	Off
Blood Suppression	Off
Wrap-up Magn.	None
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement
Reordering	Linear

Resolution - Common

FOV Read	256 mm
FOV Phase	93.8 %
Slice Thickness	0.80 mm
Base Resolution	320
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Total Factor	2
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	32
Acceleration Factor 3D	1
Phase Partial Fourier	Allowed
Slice Partial Fourier	Off
Elliptical Scanning	On

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	2D
Normalize	Prescan
Image Filter	On

Geometry - Common

Slab Group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	224
Phase Oversampling	0 %
Slice Oversampling	7.1 %
FOV Read	256 mm
FOV Phase	93.8 %
Slice Thickness	0.80 mm
TR	3200.0 ms
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm

Geometry - AutoAlign

H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	-0.80 deg

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	-0.80 deg
A >> P	240 mm
F >> H	256 mm
R >> L	180 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	5.000
Gain	High

Physio - Signal

1st Signal/Mode	None
Trigger Delay	0 ms
TR	3200.0 ms
Concatenations	1

Physio - Cardiac

Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FOV Read	256 mm
FOV Phase	93.8 %
Phase Resolution	100 %

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	spc
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Fast
Gradient Mode	Fast
Flow Compensation	None
Reordering	Linear
Bandwidth	744 Hz/Px
Echo Spacing	3.86 ms
Turbo Factor	314
Echo Train Duration	1166 ms

Sequence - Part 2

Introduction	On
--------------	----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Assistant

SAR Assistant	Off
Allowed Delay	0 s

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\ASL_ADNI

TA: 5:05 min Coil Selection: Auto Voxel Size: 3.0×3.0×4.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slices per Slab	40
Phase Oversampling	10 %
Slice Oversampling	20.0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	5000.0 ms
TE	16.60 ms
Averages	1
Concatenations	1
AutoAlign	Head > Brain

Contrast - Common

TR	5000.0 ms
TE	16.60 ms
Flip Angle	180 deg
Fat-Water Contrast	Fat Saturation
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	2
Multiple Series	Off
Delay in TR	0.00 ms
Reordering	Centric

Contrast - ASL

Perfusion Mode	FAIR Q2T
----------------	----------

Contrast - ASL

Suppression	Gray-White
Bolus Duration	800.00 ms
Inversion Time	2000.00 ms
Inversion Array Size	1

Resolution - Common

FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
Base Resolution	64
Phase Resolution	98 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off
Slice Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Prescan

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slices per Slab	40
Phase Oversampling	10 %
Slice Oversampling	20.0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	5000.0 ms
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm

Geometry - AutoAlign

H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	192 mm
R >> L	192 mm
F >> H	160 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
---------	----------

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	5000.0 ms

Physio - Signal

Segments	12
Concatenations	1

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	tgse
Dimension	3D
RF Pulse Type	Normal
Gradient Mode	Fast
Reordering	Centric
Bandwidth	2604 Hz/Px
Echo Spacing	0.49 ms
Turbo Factor	12
Segments	12
EPI Factor	23

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\QSM_3D

TA: 5:03 min Coil Selection: Auto Voxel Size: 0.9×0.9×1.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	On
Graphic segment	1st Segment
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	20 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	R >> L
Slices per Slab	128
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FOV Read	240 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
TR	44.0 ms
TE 1	3.60 ms
TE 2	9.51 ms
TE 3	15.43 ms
TE 4	21.34 ms
TE 5	27.26 ms
TE 6	33.17 ms
TE 7	39.09 ms
Averages	1
Concatenations	1
AutoAlign	Head > Brain

Contrast - Common

TR	44.0 ms
TE 1	3.60 ms
TE 2	9.51 ms
TE 3	15.43 ms
TE 4	21.34 ms
TE 5	27.26 ms
TE 6	33.17 ms
TE 7	39.09 ms
MTC	Off
Magn. Preparation	None

Contrast - Common

Flip Angle	15 deg
Fat-Water Contrast	Standard
Dark Blood	Off
Contrasts	7
SWI	Off
Reconstruction	Magn./Phase

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement

Resolution - Common

FOV Read	240 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
Base Resolution	256
Phase Resolution	100 %
Slice Resolution	50 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	24
Acceleration Factor 3D	1
Deep Resolve	Off
Phase Partial Fourier	6/8
Slice Partial Fourier	Off
Asymmetric Echo	Off
Elliptical Scanning	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	20 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	R >> L
Slices per Slab	128
Phase Oversampling	0 %
Slice Oversampling	0.0 %

Geometry - Common

FOV Read	240 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
TR	44.0 ms
Multi-Slice Mode	Interleaved
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	R >> L
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	88.17 deg

Geometry - Saturation

Saturation Mode	Standard
Special Saturation	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal

System - Adjust Volume

Rotation	88.17 deg
R >> L	240 mm
A >> P	240 mm
F >> H	128 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slab-sel.
LR Balancing	Off

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	44.0 ms
Segments	1
Concatenations	1

Physio - Cardiac

Tagging	None
Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FOV Read	240 mm
FOV Phase	100.0 %
Phase Resolution	100 %

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Liver

Liver Registration	Off
Save Original Images	On

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off

Inline - MIP

MPR Cor	Off
MPR Tra	Off

Inline - Soft Tissue

Wash-in	Off
Wash-out	Off
TTP	Off
PEI	Off
MIP Time	Off
Measurements	1

Inline - Composing

Inline Composing	Off
------------------	-----

Inline - MapIt

MapIt	T2* Map
Flip Angle	15 deg
Measurements	1
Contrasts	7
TE 1	3.60 ms
TE 2	9.51 ms
TE 3	15.43 ms
TE 4	21.34 ms
TE 5	27.26 ms
TE 6	33.17 ms
TE 7	39.09 ms
TR	44.0 ms
Noise Threshold	25
Save Original Images	On

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Asymmetric Echo	Off
Segments	1

Sequence - Part 2

Introduction	On
RF Spoiling	On
Acoustic noise reduction	Off

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Assistant

SAR Assistant	Off
Allowed Delay	0 s

Sequence - Part 1

Sequence Name	fl
Dimension	3D
Excitation	Slab-sel.
RF Pulse Type	Fast
Readout Mode	Monopolar
Gradient Mode	Fast
Flow Compensation 1	None
Flow Compensation 2	None
Flow Compensation 3	None
Flow Compensation 4	None
Flow Compensation 5	None
Flow Compensation 6	None
Flow Compensation 7	None
Bandwidth 1	240 Hz/Px
Bandwidth 2	240 Hz/Px
Bandwidth 3	240 Hz/Px
Bandwidth 4	240 Hz/Px
Bandwidth 5	240 Hz/Px
Bandwidth 6	240 Hz/Px
Bandwidth 7	240 Hz/Px

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\DWI_AP

TA: 3:29 min Coil Selection: Auto Voxel Size: 1.7×1.7×1.7 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	84
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	204 mm
FOV Phase	100.0 %
Slice Thickness	1.7 mm
TR	3600.0 ms
TE	79.00 ms
Multi-band accel. factor	3
AutoAlign	Head > Brain

Contrast - Common

TR	3600.0 ms
TE	79.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Standard
Grad. rev. fat suppr.	Enabled
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Multiple Series	Off
Delay in TR	0.00 ms

Resolution - Common

FOV Read	204 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	1.7 mm
Base Resolution	120
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	6/8

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan
Noise Masking	Off

Geometry - Common

Slice Group	1
Slices	84
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	204 mm
FOV Phase	100.0 %
Slice Thickness	1.7 mm
TR	3600.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	3

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	204 mm
R >> L	204 mm
F >> H	143 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	3600.0 ms
Multi-band accel. factor	3

Physio - PACE

Resp. Control	Off
Multi-band accel. factor	3

Diff

Diffusion Mode	Free
Diff. Directions	53
Diffusion Scheme	Monopolar
Diff. Weightings	2
b-value 1	0 s/mm ²
b-value 2	2000 s/mm ²
Averages 1	1
Averages 2	1
Dynamic Field Correction	Off
Invert Gray Scale	Off
Diff. Weighted Images	On
Trace Weighted Images	Off
Tensor	Off
FA Maps	Off
ADC Maps	Off
Exponential ADC Maps	Off
Noise Masking	Off
Calculated Image	Off

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Bandwidth	1984 Hz/Px
Echo Spacing	0.62 ms
Free Echo Spacing	Off
EPI Factor	120

Sequence - Part 2

Introduction	Off
RF Spoiling	Off
Phase Correction	Internal

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Excite pulse duration	5120 us
Refocus pulse duration	10240 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off

Sequence - Special

Time-shifted MB RF	Off
SENSE1 coil combine	On
Invert RO/PE polarity	Off
PF omits higher k-space	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\DWI_PA

TA: 3:32 min Coil Selection: Auto Voxel Size: 1.7×1.7×1.7 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	84
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	204 mm
FOV Phase	100.0 %
Slice Thickness	1.7 mm
TR	3600.0 ms
TE	79.00 ms
Multi-band accel. factor	3
AutoAlign	Head > Brain

Contrast - Common

TR	3600.0 ms
TE	79.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Standard
Grad. rev. fat suppr.	Enabled
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Multiple Series	Off
Delay in TR	0.00 ms

Resolution - Common

FOV Read	204 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	1.7 mm
Base Resolution	120
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	6/8

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan
Noise Masking	Off

Geometry - Common

Slice Group	1
Slices	84
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	204 mm
FOV Phase	100.0 %
Slice Thickness	1.7 mm
TR	3600.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	3

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	204 mm
R >> L	204 mm
F >> H	143 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	3600.0 ms
Multi-band accel. factor	3

Physio - PACE

Resp. Control	Off
Multi-band accel. factor	3

Diff

Diffusion Mode	Free
Diff. Directions	54
Diffusion Scheme	Monopolar
Diff. Weightings	2
b-value 1	0 s/mm ²
b-value 2	2000 s/mm ²
Averages 1	1
Averages 2	1
Dynamic Field Correction	Off
Invert Gray Scale	Off
Diff. Weighted Images	On
Trace Weighted Images	Off
Tensor	Off
FA Maps	Off
ADC Maps	Off
Exponential ADC Maps	Off
Noise Masking	Off
Calculated Image	Off

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Bandwidth	1984 Hz/Px
Echo Spacing	0.62 ms
Free Echo Spacing	Off
EPI Factor	120

Sequence - Part 2

Introduction	Off
RF Spoiling	Off
Phase Correction	Internal

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Excite pulse duration	5120 us
Refocus pulse duration	10240 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off

Sequence - Special

Time-shifted MB RF	Off
SENSE1 coil combine	On
Invert RO/PE polarity	On
PF omits higher k-space	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\SEField2_AP

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\BOLD_REST2_AP

TA: 5:08 min Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	34.40 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	Head > Brain

Contrast - Common

TR	800.0 ms
TE	34.40 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	375
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	375
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Flow Compensation	None
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	Off
EPI Factor	86

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg

Sequence - Special

Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\SEField2_PA

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\BOLD_REST2_PA

TA: 5:08 min Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	34.40 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	Head > Brain

Contrast - Common

TR	800.0 ms
TE	34.40 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	375
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	375
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Flow Compensation	None
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	Off
EPI Factor	86

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg

Sequence - Special

Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\SEField3_AP

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\BOLD_REST3_AP

TA: 5:08 min Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	34.40 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	Head > Brain

Contrast - Common

TR	800.0 ms
TE	34.40 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	375
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	375
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Flow Compensation	None
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	Off
EPI Factor	86

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg

Sequence - Special

Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\SEField3_PA

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\BOLD_REST3_PA

TA: 5:08 min Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	34.40 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	Head > Brain

Contrast - Common

TR	800.0 ms
TE	34.40 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	375
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	375
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Flow Compensation	None
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	Off
EPI Factor	86

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg

Sequence - Special

Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\SEField4_AP

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\SEField4_PA

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\BOLD_EMOTION_PA

TA: 4:08 min Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	34.40 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	Head > Brain

Contrast - Common

TR	800.0 ms
TE	34.40 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	300
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	300
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Flow Compensation	None
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	Off
EPI Factor	86

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg

Sequence - Special

Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\SEField5_AP

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\SEField5_PA

TA: 6 sec Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	Head > Brain

Contrast - Common

TR	6100.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	6100.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	6100.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	On
EPI Factor	86

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\USER\Brain\HARP_Prisma_XA60A\HARP_20190617\BOLD_CARIT_PA

TA: 4:08 min Coil Selection: Auto Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	34.40 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	Head > Brain

Contrast - Common

TR	800.0 ms
TE	34.40 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	300
Delay in TR	0.00 ms

Resolution - Common

FOV Read	206 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.4 mm
Base Resolution	86
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	206 mm
FOV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	-1.83 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	-1.83 deg
A >> P	206 mm
R >> L	206 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.255441 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On

BOLD

Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	300
Delay in TR	0.00 ms

Inline - Open Recon

Algorithm	None
-----------	------

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Flow Compensation	None
Bandwidth	2076 Hz/Px
Echo Spacing	0.63 ms
Free Echo Spacing	Off
EPI Factor	86

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX Delta Frequency	0 Hz
TX Nucleus	None
TX Delta Frequency	0 Hz

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg

Sequence - Special

Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----